





中国认可 国际互认 检测 TESTING **CNAS L0599**

Test Report

SL52035299246101TX

Date:September 28,2020

Page 1 of 4

THIS REPORT CANCELS AND SUPERSEDES THE TEST REPORT NO. SL52035286037801TX DATE: Sep 04, 2020 **ISSUED BY SGS (SHANGHAI) UPDATED CLIENT INFORMATION**

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description (A)Medical Face Mask Type I

SGS Internal Ref.No. 14S20015723 Sample Color (A)Blue/Pink 200801

Lot No.

Product code: SQE-1002/3 (L/M/S) Other Info.

Test Performed Selected test(s) as requested by applicant

Sample Receiving Date Aug 18, 2020

Testing Period Aug 18, 2020 - Sep 04, 2020

Unless otherwise stated the results shown in this test report refer only to the Test Result(s)

sample(s) tested, for further details, please refer to the following page(s).

Comment:

Comment.			
EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test	(A)		
Methods			
Clause 5.2 Performance Requirement			
Clause 5.2.2 Bacterial filtration efficiency (BFE)	M		
Clause 5.2.3 Breathability	M		
Clause 5.2.4 Splash Resistance	EXCLUDED		
Clause 5.2.5 Microbial Cleanliness	M		
Clause 5.2.6 Biocompatibility	EXCLUDED		

Remark: M=Meet EN 14683:2019+AC:2019 Performance Requirement (Type I) F=Below EN 14683:2019+AC:2019 Performance Requirement (Type I)

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu / Hailian Xuan (Authorized Signatory)

检验检测专用章 spection & Testing Service

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/T

3rdBuilding, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763

e sgs.china@sgs.com







中国认可 国际互认 检测 TESTING CNAS L0599

Test Report

SL52035299246101TX

Date:September 28,2020

Page 2 of 4

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial Filtration Efficiency (BFE)

(EN 14683:2019+AC:2019 Annex B)

Sample: A

Test Side : Inside

Test Area : Approximately 60 cm²

Flow Rate : 28.3 L/min

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Dimensions of test specimen : ~170mm x 170mm

Positive Control Average : 2548.5 CFU

Negative Monitor Count : <1 CFU

Mean Particle Size : 3.0 ±0.3µm

Test bacteria : Staphylococcus aureus ATCC 6538

Test Item	Specimen No.	Result
47	T.	99.9%
Bacterial Filtration Efficiency (BFE)	2	99.9%
	3 🕸	99.9%
	4	99.9%
	5	99.9%

Remark:

- 1) Performance Requirement: Type I≥95%, Type II≥98%, Type IIR ≥98%
- 2) The number of specimens that shall be tested is minimum 5, but can be greater and shall be increased if necessary to allow for an AQL(Acceptable Quality Level) of 4%.